



- NOTE:**
- 1). THE FLOWMETER MUST BE INSTALLED TO WITHIN $\pm 1^\circ$ OF THE HORIZONTAL PLANE.
 - 2). THE FLOWMETER IS BUILT TO NFPA CODE 8503 REQUIREMENTS.
 - 3). FOR FURTHER INFORMATION REFER TO THE WIRING DIAGRAM, THE COAL SCALE INSTRUCTION MANUAL OR CONTACT THE NEAREST SIEMENS MILLTRONICS REPRESENTATIVE.

1		UPDATE BORDER PER ECN 2004-0001	JLC	PTR	FEB 20/04
USE DIMENSIONS ONLY - DO NOT SCALE		REPLACES 1-7540049Z-SM-B REV 0			
DIMENSIONS ARE IN INCHES		Rev.	Revision / ECN Description	Drawn	Appr.
Third Angle Projection		Product Group	MASS DYNAMICS	Tolerance Unless Otherwise Noted: UOS	Scale:
		Date:	18 FEB	1 Place Decimal ± 0.03	Angles: $\pm 0.5^\circ$
		Drawn:	R. VASS	2 Place Decimal ± 0.01	Size: B
		Checked:		3 Place Decimal ± 0.002	
		Approved:		TITLE: 6" (152mm) C-40 COAL SCALE CUSTOMER OUTLINE DRAWING	
		Location:	PETERBOROUGH	DRAWING No: 23650450	
		SIEMENS MILLTRONICS PROCESS INSTRUMENTS INC. Peterborough, Ontario, Canada		Rev.	1
		File No.	2365045001	Plot at:	3:32
				Sheet	1 of 1

For Public Release