



- NOTES:
1. MILD STEEL OR STAINLESS CONSTRUCTION.
 2. FLOWMETER SUPPORT SHOULD BE RIGID AND INDEPENDENT OF ENCLOSURE.
 3. ANY MODIFICATION OR CHANGE SHOULD BE APPROVED BY MILLTRONICS.
 4. SEE DATA SHEET FOR MATERIAL SPECIFICATIONS.
 5. SEE INSTRUCTION MANUAL FOR LEVELING THE SENSING HEAD.
 6. () INDICATES DIMENSIONS IN MILLIMETERS.

0		FOR CONSTRUCTION		OCT 19/90
NO	SYM.	DESCRIPTION	APP.	DATE
REVISIONS				
 PETERBOROUGH, ONTARIO, CANADA.				
FILE # 10079200		DRAWING NUMBER		REV
PLOT @ 1=16		1-7510030Z-DP-B		0
JOB No. .		SHEET 1 OF 1		

CONFIDENTIAL
IMPORTANT NOTICE
 THIS DOCUMENT REMAINS THE PROPERTY OF MILLTRONICS AND IS SUBJECT TO RECALL. IT MAY NOT BE COPIED, AND IS ISSUED AND CAN BE UTILIZED ONLY FOR SUCH LIMITED PURPOSES AS MAY SPECIFICALLY HAVE BEEN AUTHORIZED BY MILLTRONICS. IT IS TO BE MAINTAINED CONFIDENTIAL, SINCE IT MAY CONTAIN PROPRIETARY INFORMATION AND TRADE SECRETS OF MILLTRONICS OR OTHERS. THE ITEMS DEPICTED MAY THEMSELVES BE THE SUBJECT OF PATENTS, INDUSTRIAL DESIGN REGISTRATIONS OR COPYRIGHTS OF MILLTRONICS OR OTHERS, AND THE ISSUE OF THIS DRAWING DOES NOT IMPLY ANY LICENSE UNDER ANY SUCH RIGHTS.
 COPYRIGHT MILLTRONICS
 ALL RIGHTS RESERVED

SCALE 1"=16"	DATE
DRAWN D.McCONNELL	OCT 19/90
CHECKED .	.
APPROVED .	.

TOLERANCES UNLESS OTHERWISE NOTED	THIRD ANGLE PROJECTION
DIMENSIONS AS A DECIMAL $\pm 0.03"$ ± 0.75 mm	ANGLE $\pm 0.5^\circ$
TITLE E-40 FLOWMETER, BASE MOUNT 2" (51) INLET OUTLINE DIMENSIONS	

1
2
3
4
5
6
7
8
9
10