



- NOTES:**
1. MILD STEEL OR STAINLESS CONSTRUCTION.
 2. FLOWMETER SUPPORT SHOULD BE RIGID AND INDEPENDENT OF ENCLOSURE.
 3. ANY MODIFICATION OR CHANGE SHOULD BE APPROVED BY MILLTRONICS.
 4. SEE DATA SHEET FOR MATERIAL SPECIFICATIONS.
 5. SEE INSTRUCTION MANUAL FOR LEVELING THE SENSING PLATE.
 6. () INDICATES DIMENSIONS IN MILLIMETERS.

0		FOR CONSTRUCTION	OCT 22/90
NO	SYM.	DESCRIPTION	APP. DATE
REVISIONS			
 PETERBOROUGH, ONTARIO, CANADA.			
FILE #	10079500	DRAWING NUMBER	REV
PLOT @	1=16	1-7510033Z-DP-B	0
JOB No.	.	SHEET 1	OF 1

<p>CONFIDENTIAL IMPORTANT NOTICE</p> <p>THIS DOCUMENT REMAINS THE PROPERTY OF MILLTRONICS AND IS SUBJECT TO RECALL. IT MAY NOT BE COPIED, AND IS ISSUED AND CAN BE UTILIZED ONLY FOR SUCH LIMITED PURPOSES AS MAY SPECIFICALLY HAVE BEEN AUTHORIZED BY MILLTRONICS. IT IS TO BE MAINTAINED CONFIDENTIAL, SINCE IT MAY CONTAIN PROPRIETARY INFORMATION AND TRADE SECRETS OF MILLTRONICS OR OTHERS. THE ITEMS DEPICTED MAY THEMSELVES BE THE SUBJECT OF PATENTS, INDUSTRIAL DESIGN REGISTRATIONS OR COPYRIGHTS OF MILLTRONICS OR OTHERS, AND THE ISSUE OF THIS DRAWING DOES NOT IMPLY ANY LICENSE UNDER ANY SUCH RIGHTS.</p> <p>COPYRIGHT MILLTRONICS ALL RIGHTS RESERVED</p>	SCALE	1"=16"	DATE		TOLERANCES UNLESS OTHERWISE NOTED	THIRD ANGLE PROJECTION		
	DRAWN	D.McCONNELL	OCT 22/90		DIMENSIONS AS A DECIMAL	±0.03" ±0.75 mm	ANGLE	± 0.5°
	CHECKED	.			TITLE	E-40 FLOWMETER, BASE MOUNT 8" (203) INLET OUTLINE DIMENSIONS		
	APPROVED	.						