



NOTES

- 1). THE FLOWMETER MUST BE INSTALLED LEVEL TO WITHIN $\pm 1^\circ$.
- 2). FOR SENSING PLATE REMOVAL, CLEARANCE OF 813 mm (32") ABOVE THE TOP COVER IS REQUIRED.
- 3). FOR FURTHER INFORMATION REFER TO THE MILLTRONICS MILLFLO INSTRUCTION MANUAL OR CONTACT YOUR SIEMENS MILLTRONICS REPRESENTATIVE.

USE DIMENSIONS ONLY - DO NOT SCALE	0	FOR RELEASE	RPC	RDC	MAR 12/04
DIMENSIONS ARE IN MILLIMETERS	Rev.	Revision / ECN Description	Drawn	Appr.	Date
Third Angle Projection 	Product Group	Tolerance Unless Otherwise Noted: UOS		Scale:	Size:
	MASS DYNAMICS	1 Place Decimal ± 0.8 Angles: $\pm 0.5^\circ$		0.0625=1	B
FOR INTERNAL USE ONLY	Date:	12 MAR 2004			
	Drawn:	R. CLYSDALE			
	Checked:	P. RICCI			
	Approved:	R. CLOSS			
	Location:	PETERBOROUGH			
SIEMENS MILLTRONICS PROCESS INSTRUMENTS INC. Peterborough, Ontario, Canada	TITLE:		250 mm (10") MILLTRONICS MILLFLO 60°, C/W DIN FLANGE CUSTOMER OUTLINE DRAWING		
File No. 2365045800	DRAWING No:		23650458		Rev. 0
Plot at: 1=16	Sheet 1		Of 1		

1
2
3
4
5
6
7
8
9
10