



NOTES:

- MILD STEEL OR STAINLESS CONSTRUCTION.
- FLOWMETER SUPPORT SHOULD BE RIGID AND INDEPENDENT OF HOUSING (CABINET)
- ANY MODIFICATION OR CHANGE SHOULD BE APPROVED BY SIEMENS MILLTRONICS.
- SEE DATA SHEET FOR MATERIAL SPECIFICATIONS.
- SEE INSTRUCTION MANUAL FOR LEVELING THE FLOWMETER SENSING HEAD.
- () INDICATES DIMENSIONS IN MILLIMETERS.

2		ADD MILLTRONICS TO PRODUCT NAME/UPDATE BORDER PER ECN 2003-0288	RPC	PTR	DEC 19/03
USE DIMENSIONS ONLY - DO NOT SCALE		-B REV 1			
DIMENSIONS ARE IN INCHES		Rev.	Revision / ECN Description	Drawn	Appr. Date
Third Angle Projection 	Product Group	Tolerance Unless Otherwise Noted: UOS		Scale:	Size:
	MASS	1 Place 2 Place Decimal ± 0.01 3 Place Decimal ± 0.002		Angles: ± 0.5°	B
FOR IN USE ONLY	Date:	25 JAN 90			
	Drawn:	TITLE: E 40 flowmeter DN150			
	Checked:	DRAWING No: 23650472			
	Approved:	Rev. 2			
	Location:	PETERBOROUGH			
SIEMENS MILLTRONICS PROCESS INSTRUMENTS INC. Peterborough, Ontario, Canada		DRAWING No: 23650472		Rev. 2	
File No.	Plot at: 1= 16		Sheet		