



NOTE

- 1). THE FLOWMETER MUST BE INSTALLED TO WITHIN $\pm 1^\circ$.
- 2). FOR SENSING PLATE REMOVAL, CLEARANCE OF 26" (660 mm) ABOVE THE TOP COVER IS REQUIRED.
- 3). FOR FURTHER INFORMATION PLEASE REFER TO THE MILLTRONICS MILLFLO INSTRUCTION MANUAL.

1		ADD MILLTRONICS TO PRODUCT NAME/UPDATE BORDER PER ECN 2003-0289	RPC	PTR	DEC 19/03
USE DIMENSIONS ONLY - DO NOT SCALE		REPLACES 1-7500055Z-SM-B	REV 0		
DIMENSIONS ARE IN INCHES		Rev.	Revision / ECN Description	Drawn	Appr. Date
Third Angle Projection 	Product Group	MASS DYNAMICS		Tolerance Unless Otherwise Noted: UOS	Scale:
	Date:	12 MAR 92		1 Place Decimal ± 0.03	0.125=1
FOR INTERNAL USE ONLY	Drawn:	R. VASS		Angles: $\pm 0.5^\circ$	Size: B
	Checked:			TITLE: MILLTRONICS MILLFLO 6" (152 mm), 60°, CUSTOMER OUTLINE DIAGRAM	
	Approved:				
	Location:	PETERBOROUGH		DRAWING No: 23650485 Rev. 1	
	SIEMENS MILLTRONICS PROCESS INSTRUMENTS INC. Peterborough, Ontario, Canada				
File No.	2365048501		Plot at: 1= 8	Sheet 1 of 1	