



NOTES

- 1) THE FLOWMETER MUST BE INSTALLED LEVEL TO WITHIN $\pm 1^\circ$.
- 2) FOR SENSING PLATE REMOVAL, CLEARANCE OF 28" (711 mm) ABOVE THE TOP COVER IS REQUIRED.
- 3) FOR FURTHER INFORMATION REFER TO THE MILLTRONICS MILLFLO INSTRUCTION MANUAL OR CONTACT THE NEAREST SIEMENS MILLTRONICS REPRESENTATIVE.

1		ADD MILLTRONICS TO PRODUCT NAME/UPDATE BORDER PER ECN 2003-0289	RPC	PTR	DEC 19/03
USE DIMENSIONS ONLY - DO NOT SCALE		REPLACES 1-7600007Z-SM-B	REV	0	
DIMENSIONS ARE IN INCHES		Rev.	Revision / ECN Description	Drawn	Appr. Date
Third Angle Projection 	Product Group	MASS DYNAMICS		Tolerance Unless Otherwise Noted:	UOS
	Date:	5 JULY 93		1 Place Decimal ± 0.03	Angles: $\pm 0.5^\circ$
FOR INTERNAL USE ONLY	Drawn:	R. VASS		2 Place Decimal ± 0.01	Scale: .0625=1
	Checked:			3 Place Decimal ± 0.002	Size: B
	Approved:			TITLE: MILLTRONICS MILLFLO, 10" (254 mm), 60°, CUSTOMER OUTLINE DIAGRAM	
	Location:	PETERBOROUGH		DRAWING No: 23650486	
	SIEMENS MILLTRONICS PROCESS INSTRUMENTS INC. Peterborough, Ontario, Canada			Rev. 1	
File No. 2365048601	Plot at: 1=16	Sheet 1 of 1			