



NOTES

- 1). THE FLOWMETER MUST BE INSTALLED LEVEL TO WITHIN $\pm 1^\circ$.
- 2). FOR SENSING PLATE REMOVAL, CLEARANCE OF 813 mm (32") ABOVE THE TOP COVER IS REQUIRED.
- 3). FOR FURTHER INFORMATION REFER TO THE MILLTRONICS MILLFLO INSTRUCTION MANUAL OR CONTACT YOUR SIEMENS MILLTRONICS REPRESENTATIVE.

USE DIMENSIONS ONLY - DO NOT SCALE		0	FOR RELEASE	RPC	RDC	MAR 12/04
DIMENSIONS ARE IN MILLIMETERS		Rev.	Revision / ECN Description	Drawn	Appr.	Date
Third Angle Projection 	Product Group	MASS DYNAMICS		Tolerance Unless Otherwise Noted: UOS		Scale:
	Date:	12 MAR 2004		1 Place Decimal ± 0.8 2 Place Decimal ± 0.3 3 Place Decimal ± 0.10		0.0625=1 B
For Public Release	Drawn:	R. CLYSDALE		TITLE: 250 mm (10") MILLTRONICS MILLFLO 60°, C/W DIN FLANGE CUSTOMER OUTLINE DRAWING		
	Checked:	P. RICCI				
	Approved:	R. CLOSS		DRAWING No:	23650458	Rev. 0
	Location:	PETERBOROUGH				
	SIEMENS MILLTRONICS PROCESS INSTRUMENTS INC. Peterborough, Ontario, Canada		Plot at:	1=16	Sheet	1 Of 1

1
2
3
4
5
6
7
8
9
10